

Compressor Nebuliser

- Piston compressor
- Power lead
- ON/OFF Switch
- Air filter compartment
- Nebuliser
- A: Advanced treatment (position A)
- S: Short treatment (position B)
- Nasal washer
- Air hose
- Mouthpiece with valve
- Nose piece
- Adult face mask
- Child face mask
- Fitting the nebuliser
- Assembling and usage of the nasal washer

Dear Customer, This nebuliser is an aerosol therapy system suitable for domestic use. This device is used for the nebulisation of liquids and liquid medication (aerosols) and for the treatment of the upper and lower respiratory tract. If you have any questions, problems or want to order spare parts please contact your local MicroLife-Customer Service. Your dealer or pharmacist will be able to give you the address of the MicroLife dealer in your country. Alternatively, visit the internet at www.microlife.com where you will find a wealth of invaluable information on our products. Stay healthy – MicroLife AG!

1. Explanation of Symbols

- Electronic devices must be disposed of in accordance with the local applicable regulations, not with domestic waste.
- Read the instructions carefully before using the device.
- Type BF applied part
- Class II equipment
- Serial number
- Reference number
- Manufacturer
- ON
- OFF

- IP21 Protection against solid foreign objects and harmful effects due to the ingress of water
- CE Marking of Conformity
- Important Safety Instructions
- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

2. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

3. Preparation and Usage of This Device

- Assembly of the nebuliser kit. Ensure that all parts are complete.
- Fit the nebuliser with the inhalation solution as per your doctor's instructions. Ensure that you do not exceed the maximum level.
- Connect the nebuliser with the air hose (2) to the compressor (1) and plug the power lead (3) into the socket (230V 50 Hz AC).
- Start the treatment, set ON/OFF switch (3) into the «A» position.
- The mouthpiece gives you a better drug delivery to the lungs.
- Choose between adult (2) or child face mask (3) and make sure that it encloses the mouth and nose as completely as possible.
- Use all accessories including the nose piece (4) as prescribed by your doctor.
- During inhalation, sit upright and relaxed at a table and not in an armchair, in order to avoid compressing your respiratory airways and impairing the therapeutic effectiveness. Do not lie down while inhaling. Stop inhalation if you feel unwell.

4. NEB 2 in 1 - Aerosol therapy

- The advanced NEB 2 in 1 aerosol therapy allows effective treatments of the medium and lower respiratory tract (asthma, chronic bronchitis, cystic fibrosis etc.).
- Advanced treatment**
- This treatment option allows the maximum amount of medication to be delivered thus reaching the respiratory capacity of each patient. The valve system optimally adapts the flow of medication during inhalation, reducing resistance during expiration.
- Use the breath-enhanced valve system (synchronized with the patient's breathing) to get the maximal benefit.
- Close the valve on the nebuliser top (2-A) (position A) and follow the steps described in Section 3.2.
- Fill with medication (3) and connect the nebuliser top to the nebuliser bottom by rotating it clockwise.
- Follow the usage steps as described in Section 3.4.
- To enable the synchronized breathing function of the nebuliser, use only the mouthpiece with the expiration valve (3). Do not use with masks or nose piece.
- Short treatment**
- The therapeutic session will be reduced if the valve is open (position B) (3-B).
- A synchronized nebulisation is not possible using this treatment option.

5. Nasal washer

- The nasal washer (6) is a medical device suitable for cleaning nasal cavities. The nasal washer is used to clean the nasal mucosa and creates a mucociliary jet for the treatment of respiratory problems for the upper airways (allergic and non-allergic rhinitis, rhinosinusitis, polyposis, etc.).
1. Remove the nasal washer cover (3-B) by rotating counter clockwise.
2. Fill the container of the nasal washer with isotonic solution (3-B, 2). Ensure that you do not exceed the maximum level.
3. Use the cover tightly by rotating clockwise (3-B, 3).
4. Connect the nasal washer with the air hose (3-B, 4) to the compressor (1) and plug the power lead (3) into the socket (230V 50 Hz AC).
5. Close the non-treated nostril with your finger.
6. Position the nasal washer cover on nostril and inhale through the nose (3).
7. Switch the ON/OFF switch (3) to position «A» to turn the device off and place your finger over the aperture on the nasal washer to start the treatment (3).
- We recommend changing from one nostril to the other every 10-20 seconds.
- Children and people in need of care must be assisted for use.
- Only use with isotonic saline solution.

6. Cleaning and Disinfecting

- Thoroughly clean all components to remove medication residuals and possible impurities after each treatment. Use a soft and dry cloth with non-abrasive cleaners to clean the device.
- Make sure that the internal parts of the device are not in contact with liquids and that the power plug is disconnected.
- Cleaning and disinfecting of the accessories**
- Follow carefully the cleaning and disinfecting instructions for each accessory.
- Do not use the device for cleaning or disinfecting of the device and the success of the therapy.
- Before and after each treatment**
- Turn the device off by turning the top counter-clockwise and remove the nebulisation cone. Wash the components of the disassembled nebuliser, the mouthpiece and the nasal washer with warm water (30-35 °C) in boiling water for 5 minutes. Reassemble the nebuliser components and connect to the air hose connector, switch the device on and use it as described in the "Technical Specifications" section.
- Wash the air tube with warm water.
- Only use cold disinfecting liquids following the manufacturer's instructions.
- Do not boil nor autoclave the air tube.

7. Maintenance, Care, and Service

- Order all spare parts from your dealer or pharmacist, or order them directly from the manufacturer (see previous section).
- Replacement of the nebuliser**
- Replace the nebuliser (3) after a long period of inactivity, in cases where it shows deformations, breakage, or when the power lead (3-B) is damaged or the power plug is not working. We recommend to replace the nebuliser after a period between 6 months and 1 year depending on the usage.
- Replacement of the air filter**
- In normal conditions of use, the air filter (4) must be replaced approximately after 500 hours working or after 6 months of use, depending on the use.
- The filter (10-12 treatments) and if the filter shows a grey or brown color (it is wet, replace it). Extract the filter and replace it with a new one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

8. Malfunctions and Actions to take

- The device cannot be switched on.
- Ensure the power lead (3) is correctly plugged into the power socket.
 - Ensure the ON/OFF switch (3) is in the position «A».
- The nebuliser functions poorly or not at all.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating. - Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
 - Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
 - Do not use the device for any other purpose than the treatment that is recommended by your doctor.
 - Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
 - Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
 - Do not position the equipment so that it is difficult to operate the disconnection device.
 - For greater hygienic safety, do not use the same accessories for more than one person.
 - Never bend the nebuliser over 60°.
 - Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:

9. Technical Specifications

- Nebulisation rate: 0.37 ml/min. (NaCl 0.9%)
- Particle size: 0.7 µm (NaCl 0.9%)
- Particle size: 3.5 µm (MMAD)
- Max. free air flow: 15 l/min.
- Operating air flow: 5.31 l/min.
- Acoustic noise level: 52 dBA
- Power source: 230V 50 Hz AC
- Power lead length: 1.6 m
- Nebuliser capacity: min. 2 ml; max. 8 ml
- Residual volume: 0.8 ml
- Operating conditions: 10 °C - 40 °C / 50 °F - 104 °F
- Operating conditions: 30 - 85 % relative maximum humidity
- Storage and shipping: 20 °C - 60 °C / 68 °F - 140 °F
- conditions: 30 - 85 % relative maximum humidity
- Weight: 1765 g
- Dimensions: 289 x 180 x 127 mm
- IP Class: IP21
- Reference aux normes: EN13544-1; EN60601-1; EN60601-1-1
- Esperance de fonction: 1000 hours
- Capacité du réservoir: min. 2 ml; max. 8 ml
- Volume résiduel: 0.8 ml
- Limites de fonctionnement: 10 °C - 40 °C / 50 °F - 104 °F
- Humidité relative: 30 - 85 % max.
- Conditions d'utilisation: 20 °C - 60 °C / 68 °F - 140 °F
- Humidité relative: 30 - 85 % max.
- Conditions de stockage et de transport: 20 °C - 60 °C / 68 °F - 140 °F
- Humidité relative: 30 - 85 % max.
- Poids: 1765 g
- Dimensions: 289 x 180 x 127 mm
- Classe IP: IP21
- Références aux normes: EN13544-1; EN60601-1; EN60601-1-1
- Esperance de fonction: 1000 hours

10. Technical Specifications

- Nebulisation rate: 0.37 ml/min. (NaCl 0.9%)
- Particle size: 0.7 µm (NaCl 0.9%)
- Particle size: 3.5 µm (MMAD)
- Max. free air flow: 15 l/min.
- Operating air flow: 5.31 l/min.
- Acoustic noise level: 52 dBA
- Power source: 230V 50 Hz AC
- Power lead length: 1.6 m
- Nebuliser capacity: min. 2 ml; max. 8 ml
- Residual volume: 0.8 ml
- Operating conditions: 10 °C - 40 °C / 50 °F - 104 °F
- Operating conditions: 30 - 85 % relative maximum humidity
- Storage and shipping: 20 °C - 60 °C / 68 °F - 140 °F
- conditions: 30 - 85 % relative maximum humidity
- Weight: 1765 g
- Dimensions: 289 x 180 x 127 mm
- IP Class: IP21
- Reference aux normes: EN13544-1; EN60601-1; EN60601-1-1
- Esperance de fonction: 1000 hours
- Capacité du réservoir: min. 2 ml; max. 8 ml
- Volume résiduel: 0.8 ml
- Limites de fonctionnement: 10 °C - 40 °C / 50 °F - 104 °F
- Humidité relative: 30 - 85 % max.
- Conditions d'utilisation: 20 °C - 60 °C / 68 °F - 140 °F
- Humidité relative: 30 - 85 % max.
- Conditions de stockage et de transport: 20 °C - 60 °C / 68 °F - 140 °F
- Humidité relative: 30 - 85 % max.
- Poids: 1765 g
- Dimensions: 289 x 180 x 127 mm
- Classe IP: IP21
- Références aux normes: EN13544-1; EN60601-1; EN60601-1-1
- Esperance de fonction: 1000 hours

11. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

12. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

13. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

14. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

15. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.
- Do not use the device for any other purpose than the treatment that is recommended by your doctor.
- Only use the power plug if expressly indicated by your doctor. Do not use the device if the power plug is damaged or if the burflations into the nose, but only positioning them as close as possible.
- Check in the medicine instruction leaflet for possible contraindications for use with common aerosol therapy systems.
- Do not position the equipment so that it is difficult to operate the disconnection device.
- For greater hygienic safety, do not use the same accessories for more than one person.
- Never bend the nebuliser over 60°.
- Do not use this device close to any electromagnetic fields such as mobile telephones or radio installations. Keep a minimum distance of 3.3 m from such devices. When using this device:
- Ensure that children do not use this device unsupervised, some parts are small enough to be swallowed. Be aware of the risk of strangulation in case this device is supplied with cables or tubes.

16. Important Safety Instructions

- Follow instructions for use. This document provides important product operation and safety information regarding this device. Please read the instructions thoroughly before using the device and keep for future reference.
- This device may only be used for the purposes described in these instructions. The manufacturer cannot be held liable for damage caused by incorrect application.
- Do not operate the unit in presence of any anesthetic mixture inflammable with oxygen or nitrogen provokes.
- This device is not suitable for anaesthesia and lung ventilation.
- This device should only be used with original accessories as shown in these instructions.
- Do not use this device if the unit is damaged or notice anything unusual.
- Never open the device.
- This device comprises sensitive components and must be treated with caution. Observe the storage and operating conditions described in the "Technical Specifications" section.
- Protect it from:
- water and moisture
 - extreme temperatures
 - impact and dropping
 - contamination and dust
 - direct sunlight
 - heat and cold
- Comply with the safety regulations concerning the electrical devices and in particular:
- Never touch the device with wet or moist hands.
 - Place the device on a stable surface and avoid use during its operation.
 - Do not pull the power cord or the device itself to plug it from a power socket.
 - The power plug is a separate element from the grid power, keep the plug accessible when the device is in use.
- Before plugging in the device, make sure that the electrical rating, shown on the rating plate on the bottom of the device, corresponds to the electrical rating in the case of the power plug provided with the device does not fit your wall socket. In this case, contact your dealer for a replacement plug with that of a suitable one.
- Do not try to clean the filter for reuse it.
- In general, the use of adapters, simple or multiple, and/or extension cables is not recommended. If their use is indispensable, it is necessary to use types complying with safety regulations, paying attention that they do not exceed the maximum power, indicated on the adapters and extension cables.
- Do not leave the unit plugged in when not in use. Unplug the device from the wall socket when it is not in operation.
- The installation must be carried out according to the instructions of the manufacturer. An improper installation can cause damage to persons, animals or objects, for which the manufacturer cannot be held responsible.
- Do not replace the parts of the device. In case of a power lead damage, contact a technical service center authorized by the manufacturer for the replacement. The power supply cord should always be fully unwound in order to prevent dangerous overheating.
- Before performing any repair or servicing operation, turn off the device and disconnect the plug from the main supply.
- Only use the medication prescribed by your doctor and follow your doctor's instructions with regard to dosage, duration and frequency of the therapy.</

